

Creatinina

I.D.: 49561721

Data publicarii	07.09.20	Coduri CPV	33696500-0
-----------------	----------	------------	------------

Pretul estimativ: 72,22 RON - 72,22 RON

Descriere: COMPOZITIE A.Reactiv. Acid picric 25 mmol/L B.Reactiv . Hidroxid de sodiu 0.4 mol/L, detergent S.Glucoza/uree/creatinina standard. Glucoza 100 mg/dl,ureea 50 mg/dl, creatinina 2 mg/dl=177 μ mol/l. Standard lichid primar ECHIPAMENT ADITIONAL •Analizor, spectrofotometru sau fotometru cu filtru de 500 ± 20 nm., si cuva termostatabila la 37°C. •Baie termostata la 37 oC PROBA Ser, plasma sau urina, recoltate prin proceduri standard. CARACTERISTICI METROLOGICE •Limita de detectie: 0.03 mg/dl = 2.65 μ mol/l creatinina •Limita linearitatii: 20 mg/dl = 1768 μ mol/l creatinina. Pentru valori mai mari diluati proba 1/2 cu apa distilata si repetati masurarea •Repetabilitate (in cadrul unei serii de masurari) Concentratie medie CV n 1.7 mg/dl=150 μ mol/l 2.9% 20 5.3mg/dl = 468 μ mol/l 1.3% 20 •Reproductibilitate(de la o serie de masurari la alta). Concentratie medie CV n 1.7 mg/dl=150 μ mol/l 3.9% 25 5.3mg/dl = 468 μ mol/l 2.9% 25 •Sensibilitate: 31 mA*dl/mg = 0.351mA*/ μ mol. •Veridicitate: rezu
