

Estimare pret Aparat LC MS/MS

I.D.: 18508409

Data publicarii	22.03.17	Coduri CPV	39300000
-----------------	----------	------------	----------

Termenul limita pentru depunere:	24.03.17
----------------------------------	----------

Descriere: POTRIVIT PENTRU EFECTUAREA ANALIZELOR DE DETERMINARE :-COLORANTI ROSU SUDAN - COLORANTI ALIMENTARI,-INDULCITORI Specificatii ANALIZOR LC-MS/MS - pentru determinarea colorantilor ilegali - SURSA: Elsevier, Science Direct, Determination of 20 synthetic dyes in chili powders and syrup-preserved fruits by liquid chromatography/tandem mass spectrometry Analizor LC system coupled with a triple quadrupole mass spectrometer with a Turbo V ion source si Software UHPLC Solvent Delivery Smart Flow ensures optimal performance independent of mobile phase composition Spin Flow mixing technology perfectly balances gradient delay volume against mobile phase mixing efficiency Pump Type - Binary Gradient Fast LC-MS Sample Handling Reliable, precise, and accurate injections Cycle times as low as 15 s Autosampler Type - Thermostatting (complet echipat) Min 4°C Vialuri brune 1,5mL - 100buc Column Compartment - Thermostatting Holds up to 4 columns Electrospray ionization mass spectra - ESI-MS ESI+ ESI- High performance High sensitivity High selectivity High recovery Low LOQ Dwell time 0.01 Infuzie MS Instruire 3 persoane Coloana cromatografica Acclaim Polar Advantage C16 (3um, 120 A, 4.6x150mm) - 2 buc (sau echivalent) Nitrogen - nitrogen generator inside laboratory + sistem de cuplare + gaz fragmentare (butelie) Calificare operationala UPS asigura continuitatea alimentarii cu energie electrica Biblioteca spectre Posibilitate exportare/copiere date din software in excel Membrane filtrante nylon 0.45um - 100buc Acetat de amoniu 100g Pompe pentru vacuum Cele necesare pentru instalarea, punerea in functiune si efectuarea calificarii operationale si a instruirii personalului.
