
REPARATIE MONITOR FV CONTEC CMS 8000 - Chirurgie Oncologica

I.D.: 35294952

Data publicarii	15.04.19	Coduri CPV	50800000-3
-----------------	----------	------------	------------

Pretul estimativ: 2.347,90 RON - 2.347,90 RON

Descriere: Constatate defect * Depanare modul ECG * Depanare insuflatie aer (inlocuire sensor) * Resetare meniuri ,TESTARE CU SIMULATOR PACIENT! PRELUARE,DEPANARE,RETUR,PROBE FUNCTIONALE. Prezenta tine loc de comanda ferma.
