

## Echipamente de laborator- Ph metru

I.D.: 61454482

---

Data publicarii	29.09.21	Coduri CPV	38430000-8
-----------------	----------	------------	------------

---

Pretul estimativ: 18.740,00 RON - 18.740,00 RON

---

Descriere: pH-metru Model: SevenCompact Duo S213-Standard kit Producator: Mettler Toledo; Domeniu de masura pH: -2.000 ... 20.000; Rezolutie pH: 0.001/0.01/0.1; Acuratete pH:  $\pm 0.002$ ; Domeniu de masura mV: -2.000 ... 2000.0; Rezolutie mV: 0.1/1; Acuratete mV:  $\pm 0.1$ mV in-1000...1000mV;  $\pm 0.2$ mV; Domeniu de masura mV rel.: -2.000 ... 2000.0; Rezolutie mV rel. : 0.1/1; Acuratete mV rel:  $\pm 0.1$ mV in-1000...1000mV;  $\pm 0.2$ mV; Domeniu de masura al conductivitatii: 0.001  $\mu$ S/cm...1000mS/cm; Rezolutie conductivitate: 0.001 ... 1; Acuratete conductivitate:  $\pm 0.5\%$ ; Domeniu de masura TDS: 0.00mg/L...1000 g/L; Rezolutie TDS: 0.01...1; Acuratete TDS:  $\pm 0.5\%$ ; Domeniu de masura al salinitatii: 0.00...80.00 psu/ppt; Rezolutie salinitate:0.00...1.0; Acuratete salinitate:  $\pm 0.5\%$ ; Domeniu de masura rezistivitate: 0.00...100.0 M $\Omega$ •cm; Rezolutie rezistivitate: 0.00...1.0 ; Acuratete rezistivitate:  $\pm 0.5\%$ ; Domeniu de masura al conductivitatii de cenusa: 0.000...2022%; Rezolutie conductivitate cenusa: 0.001...1; Acuratete conductivitate cenusa:  $\pm 0.5\%$ ; Domeniu de masura al temperaturii MTC/ATC: -5.0 (-30.0)...130.0°C (MTC)ATC; Rezolutie temperatura MTC/ATC: 0.1°C; Acuratete temperatura MTC/ATC:  $\pm 0.1$ °C; Calibrare: Puncte de clairare; Conductivitatea 1; pH Max 5; Livrarea include: - agitator magnetic uMix (cod 30040000); - electrod de pH InLab Expert Pro-ISM; - electrod de conductivitate InLab® 731; - cate 2 plicuri cu solutii de calibrare de 4.01, 7.00, 9.21, 10.00 - cate 2 plicuri cu solutii de calibrare de 1433  $\mu$ S/cm 2 buc

---