



## Bilirubina Directa

I.D.: 56137155

---

Data publicarii	15.04.21	Coduri CPV	33696200-7 33696500-0
-----------------	----------	------------	-----------------------

---

Pretul estimativ: 81,93 RON - 81,93 RON

---

Descriere: COMPOZITIE AT. Reactiv. Acid sulfanilic 29 mmol/l, acid clorhidric 0.2 mol/l, cetrimide 50 mmol/l AD. Reactiv. Acid sulfanilic 35 mmol/l, acid clorhidric 0.24 mol/l B. Reactiv. Azotat de sodiu 11.6 mmol/l ECHIPAMENT •spectrofotometru cu filtru de 540nm(520-560) si cuva de 1 cm termostatabila la 37°C. Ser recoltat prin proceduri standard. CARACTERISTICI •Limita de detectie(bilirubina totala): 0.03 mg/dl = 0.51 µmol/l •Limita de detectie(bilirubina directa): 0.05 mg/dl = 0.85 µmol/l •Limita linearitatii: 15 mg/dl = 257 µmol/l Pentru valori mai mari diluati proba 1/3 cu apa distilata si repetati masurarea •Repetabilitate ( in cadrul unei serii de masurari)-bilirubina totala: Concentratie medie CV n 0.59 mg/dl = 10.1 µmol/l 3.0 % 20 6.74 mg/dl = 115.2 µmol/l 1.0% 20 •Repetabilitate ( in cadrul unei serii de masurari)-bilirubina directa: Concentratie medie CV n 0.81 mg/dl = 13.8 µmol/l 2.9 % 20 1.41 mg/dl = 24.1 µmol/l 2.4% 20 •Reproductibilitate( de la o serie de masurari la alta)-bilirubina totala: Concentratie medie CV n 0.59 mg/dl = 10.1 µmol/l 3.6 % 25 6.74 mg/dl = 115.2 µmol/l 3.3 % 25 •Reproductibilitate( de la o serie de masurari la alta)-bilirubina directa: Concentratie medie CV n 0.81 mg/dl = 13.8 µmol/l 7.1 % 25 1.41 mg/dl = 24.1 µmol/l 6.3 % 25 •Sensibilitate(bilirubina totala): 88 mA\*dl/mg = 5.15 mA\*l/µmol •Sensibilitate(bilirubina directa): 48 mA\*dl/mg = 2.81 mA\*l/µmol

---