



Bilirubina Totala si Directa

I.D.: 39842312

Data publicarii	06.06.18	Coduri CPV	33696200-7
-----------------	----------	------------	------------

Pretul estimativ: 64,82 RON - 64,82 RON

Descriere: Bilirubina Totala si Directa COMPOZITIE AT. Reactiv. Acid sulfanilic 29 mmol/l, acid clorhidric 0.2 mol/l, cetrimide 50 mmol/l AD. Reactiv. Acid sulfanilic 35 mmol/l, acid clorhidric 0.24 mol/l B. Reactiv. Azotat de sodiu 11.6 mmol/l ECHIPAMENT ADITIONAL • Analizor, spectrofotometru sau fotometru cu filtru de 540nm(520-560) si cuva de 1 cm termostatabila la 37°C. PROBA Ser recoltat prin proceduri standard. CARACTERISTICI METROLOGICE • Limita de detectie(bilirubina totala): 0.03 mg/dl = 0.51 µmol/l • Limita de detectie(bilirubina directa): 0.05 mg/dl = 0.85 µmol/l • Limita linearitatii: 15 mg/dl = 257 µmol/l Pentru valori mai mari diluati proba 1/3 cu apa distilata si repetati masurarea • Repetabilitate (in cadrul unei serii de masurari)-bilirubina totala: Concentratie medie CV n 0.59 mg/dl = 10.1 µmol/l 3.0 % 20 6.74 mg/dl = 115.2 µmol/l 1.0% 20 • Repetabilitate (in cadrul unei serii de masurari)-bilirubina directa: Concentratie medie CV n 0.81 mg/dl = 13.8 µmol/l 2.9 % 20 1.41 mg/dl = 24.1 µmol/l 2.4% 20 • Reproductibilitate(de la o serie de masurari la alta)-bilirubina totala: Concentratie medie CV n 0.59 mg/dl = 10.1 µmol/l 3.6 % 25 6.74 mg/dl = 115.2 µmol/l 3.3 % 25 • Reproductibilitate(de la o serie de masurari la alta)-bilirubina directa: Concentratie medie CV n 0.81 mg/dl = 13.8 µmol/l 7.1 % 25 1.41 mg/dl = 24.1 µmol/l 6.3 % 25 • Sensibilitate(bilirubina totala): 88 mA*dl/mg = 5.15 mA*/µmol • Sensibilitate(bilirubina directa): 48 mA*dl/mg = 2.81 mA*/µmol m