
Spectrophotometer Standards F0 F1 F2 F3 F4 Cert

I.D.: 97976039

Data publicarii	28.05.24	Coduri CPV	34913000-0
-----------------	----------	------------	------------

Pretul estimativ: 9.037,00 RON - 9.037,00 RON

Descriere: Nr. Denumire componenta Cod Pret Lei/buc 1 Spectrophotometer Standards F0 F1 F2 F3 F4 Cert Reference Material Standard Set for testing the wavelength accuracy and the photometric accuracy, consisting of: F1: Holmium Oxide Glass Filter, F2: Neutral Density Glass Filter (0.25 Abs), F3: Neutral Density Glass Filter (0.5 Abs), F4: Neutral Density Glass Filter (1.0 Abs), F0: Reference - empty reference material mount Wavelength accuracy certified at: wavelengths: 279, 361, 453, 536, 638nm slit width: 1nm Photometric accuracy certified at: wavelengths: 440, 465, 546.1, 590, 635nm slit width: 1nm with serial number In high-quality storage case with DAkkS calibration certificates 634-9691 9037 VALOARE TOTALA FARA TVA 9037
