
Reparatie electrocauter Autocon III 300

I.D.: 66389233

Data publicarii	16.02.22	Coduri CPV	50422000-9
-----------------	----------	------------	------------

Pretul estimativ: 2.430,00 RON - 2.430,00 RON

Descriere: Reparatie electrocauter Autocon III 300 DEFECTIUNI CONSTATATE priza conectare cablu neutral defecta, contact electric intermitent OPERATIUNI EXECUTATE demontare unitate electrochirurgie.Inlocuire priza electrod neutral, verificare touch screen, efectuat verificari conform manual de service. Efectuat actualizare soft.Efectuat testare functionalitate si electrosecuritate
