
URINALYSIS Control Positive

I.D.: 55629800

Data publicarii	31.03.21	Coduri CPV	33696500-0
-----------------	----------	------------	------------

Pretul estimativ: 184,80 RON - 184,80 RON

Descriere: URINALYSIS Control Positive 1x8ml DIRUI INDUSTRIAL CO. LTD PREZENTARE: Lichid, valabilitate 12 luni la +2-8C sigilat si protejat de lumina soarelui; dupa deschidere valabilitate o luna la +2-8C sigilat si protejat de lumina soarelui. ANALITI: Glucoza, Bilirubina, Cetone, densitate, pH, sange, Proteine, Urobilinogen, Nitriti, Leucocite, Microalbumina, Creatinina si Calciu. Valori pentru analizoarele: DIRUI (A10,E10,M10,H series,H800), BAYER(10SG), ROCHE(Combur10), KYOTO(10EA)
